



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/784,385	02/23/2004	Daniel Seguin	MOSA-09	8425
26875	7590	10/10/2006	EXAMINER	
WOOD, HERRON & EVANS, LLP 2700 CAREW TOWER 441 VINE STREET CINCINNATI, OH 45202			DUNWOODY, AARON M	
			ART UNIT	PAPER NUMBER
			3679	

DATE MAILED: 10/10/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

<b>Office Action Summary</b>	<b>Application No.</b>	<b>Applicant(s)</b>	
	10/784,385	SEGUIN ET AL.	
	<b>Examiner</b>	<b>Art Unit</b>	
	Aaron M. Dunwoody	3679	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

#### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

#### Status

- 1) ☒ Responsive to communication(s) filed on 15 August 2005.
- 2a) ☒ This action is **FINAL**.                      2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

#### Disposition of Claims

- 4) ☒ Claim(s) 1-5 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-5 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

#### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

#### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- \* See the attached detailed Office action for a list of the certified copies not received.

#### Attachment(s)

- |  |   |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)                     | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____                                      |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)          | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____  | 6) <input type="checkbox"/> Other: _____                          |

**DETAILED ACTION**

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-5 are rejected under 35 U.S.C. 103(a) as being unpatentable over US patent 4257268, Pepicelli et al in view of US patent 6117394, Smith.

In regards to claim 1, Pipicelli et al disclose a mounting fixture for connecting pipettes of various sizes to the air channel of a pipettor, the fixture comprising a tubular nose (240) having a proximate end secured to the pipettor and an open distal end; and a resilient retainer (242) lining the interior of the nose, the retainer having an axial passageway communicating at its entry and exit openings respectively with the open distal end of the nose and with the air channel of the pipettor, the passageway being tapered from a maximum diameter at the entry opening to a minimum diameter at the exit opening, the entry opening being sized to axially receive the largest of the pipettes and the sleeve being internally configured to resiliently grip differently sized pipettes inserted therein at different locations along the length thereof. Pepicelli does not disclose a stabilizing member at the open distal end of the nose. Smith teaches a stabilizing member (Tp) at the open distal end of the nose (P) to limit the possibility of cross contamination between samples (col. 1, lines 8-13). As Smith relates to disposable tips for pipette devices, it would have been obvious to one having ordinary

Art Unit: 3679

skill in the art at the time the invention was made to provide a stabilizing member at the open distal end of the nose to limit the possibility of cross contamination between samples, as taught by Smith.

In regards to claim 2, Smith further discloses the stabilizing member comprising a tubular sleeve projecting axially from the distal end of the nose.

In regards to claim 3, Smith further discloses the stabilizing member being rigid, and being detachably secured to the distal end of the nose, and having an inner diameter smaller than that of the entry opening of the resilient retainer.

In regards to claim 4, Smith further discloses the stabilizing member being snap fitted into the open distal end of the tubular nose.

In regards to claim 5, Pepicelli et al in view of Smith disclose the claimed invention except for the stabilizing member projecting axially from the distal end of the nose by a distance at least about 0.5 times the minimum diameter of the axial passageway in the resilient retainer. It would have been obvious to one having ordinary skill in the art at the time the invention was made to fabricate the stabilizing member projecting axially from the distal end of the nose by a distance at least about 0.5 times the minimum diameter of the axial passageway in the resilient retainer, since it has been held that discovering an optimum value of a result effective variable involves only routine skill in the art. *In re Boesch*, 617 F.2d 272, 205 USPQ 215 (CCPA 1980).

***Response to Arguments***

Applicant's arguments filed 8/15/2005 have been fully considered but they are not persuasive.

In response to applicant's argument that there is no suggestion to combine the references, the examiner recognizes that obviousness can only be established by combining or modifying the teachings of the prior art to produce the claimed invention where there is some teaching, suggestion, or motivation to do so found either in the references themselves or in the knowledge generally available to one of ordinary skill in the art. See *In re Fine*, 837 F.2d 1071, 5 USPQ2d 1596 (Fed. Cir. 1988) and *In re Jones*, 958 F.2d 347, 21 USPQ2d 1941 (Fed. Cir. 1992). In this case, as Smith relates to disposable tips for pipette devices, it would have been obvious to one having ordinary skill in the art at the time the invention was made to provide a stabilizing member at the open distal end of the nose to limit the possibility of cross contamination between samples, as taught by Smith.

In response to applicant's argument that the examiner's conclusion of obviousness is based upon improper hindsight reasoning, it must be recognized that any judgment on obviousness is in a sense necessarily a reconstruction based upon hindsight reasoning. But so long as it takes into account only knowledge which was within the level of ordinary skill at the time the claimed invention was made, and does not include knowledge gleaned only from the applicant's disclosure, such a reconstruction is proper. See *In re McLaughlin*, 443 F.2d 1392, 170 USPQ 209 (CCPA 1971).

***Conclusion***

**THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Aaron M. Dunwoody whose telephone number is 571-272-7080. The examiner can normally be reached on 7:30 am - 4:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Daniel P. Stodola can be reached on 571-272-7087. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3679

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Aaron M Dunwoody  
Primary Examiner  
Art Unit 3679

.amd